



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Temperature Product	Module Sensor	Temperature	74	°F
Model	MS - T&H - 101	Rel. Humidity	30	%
SN	25J-00515	Bar. Pressure	28.8	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument	
			Min	Max		
Cal Lab Probe & Test Module	Spec					
Temperature (°F)	1	75.2	-0.3	+0.3	75.4	
	2	205.9	-2.6	+2.6	206.3	
	3	-5.3	-1.6	+1.6	-5.2	

Indicates out of tolerance condition -----↑

Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	11-Aug-25	11-Aug-27
Temperature	21396189	24-Feb-25	24-Feb-27

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

 Calibrated By

10-Dec-2025 10-Dec-2027

 Calibration Date Date Due



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Pressure / Velocity Module	Temperature	75	°F
Model	S-PVF-1	Rel. Humidity	30	%
SN	25D-01113	Bar. Pressure	28.8	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1	
	1	20.0			20.0
	2	28.9			28.8
	3	33.0			33.1
Differential Pressure (in wc)	Spec		-2%-.001	+2%+.001	
	1	10.00			9.947
	2	2.000			1.990
	3	0.5000			0.4946
	4	0.0500			0.0501
	5	-10.00			-9.981
Via Pitot >> Velocity Pressure >> (inW.C. / FPM) -3% -7	6	-0.0500			-0.0495
			-3% - 7	+3% + 7	
	7	0.00065 / 102			103
	8	0.016 / 506			504

Indicates out of tolerance condition -----↑

NIST-Traceable Lab Calibration Standards

Variable	System ID	Calibration Last	Calibration Due
Pressure	7481227 / 7568470	8-Mar-23	28-Dec-25
Pressure	7871917 / 7870754	12-Sep-23	28-Dec-25
Pressure	11269047 / 12238595	25-Jul-24	25-Jul-26
Pressure	2205000006	10-Oct-25	10-Oct-27
Pressure	1208000080	14-Aug-25	14-Aug-27
Pressure	41001F6C	19-Jun-24	19-Jun-26
Pressure	41001O9L	28-Apr-25	28-Apr-27
Velocity	2500067	10-Sep-25	10-Sep-27
Velocity	2100190A	11-Aug-25	11-Aug-27

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

 Calibrated By

9-Dec-2025 9-Dec-2027

 Calibration Date Date Due



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Humidity Sensor	Temperature	75	°F
Model	PR-TH-12	Rel. Humidity	30	%
SN	2500156	Bar. Pressure	28.8	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Temperature (°F)	Spec				
	1	40.2	-1.0	1.0	39.6
	2	75.3	-1.0	1.0	75.1
	3	86.3	-1.0	1.0	85.8
	4	128.6	-2.0	2.0	128.8
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1	
	1	20.0			20.0
	2	28.7			28.7
	3	33.0			33.0
Humidity %RH 10 to 90%	Spec		-3	3	
	1	9.5			10.3
	2	25.0			25.8
	3	60.1			59.5
	4	93.0			92.3

Indicates out of tolerance condition -----↑

Calibration Standard

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	11-Aug-25	11-Aug-27
Temperature	21396189	24-Feb-25	24-Feb-27
Pressure	2205000006	10-Oct-25	10-Oct-27
Pressure	1208000080	14-Aug-25	14-Aug-27
Humidity	20558772	8-Oct-25	8-Oct-26
Humidity	20052171	13-Feb-25	13-Feb-26

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

Temperature accuracy (dry bulb) varies across the operating range:

Temperature over 32-100F +/- 1.0 F
 Temperature over 100-158F +/- 2.0 F

 Calibrated By

3-Dec-2025 3-Dec-2026

 Calibration Date Date Due



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry			Calibration Environment		
Temperature Product		Module	Probe	Temperature	75	°F
Model			PR-T-4-6	Rel. Humidity	30	%
SN			2500590	Bar. Pressure	28.8	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument	
			Min	Max		
Cal Lab Module & Test Probe	Spec					
Temperature (°F)	1	75.0	-0.3	+0.3	74.8	
	2	206.6	-2.6	+2.6	207.1	
	3	-5.2	-1.6	+1.6	-5.5	

Indicates out of tolerance condition -----↑

Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	11-Aug-25	11-Aug-27
Temperature	21396189	24-Feb-25	24-Feb-27

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

 Calibrated By

9-Dec-2025 9-Dec-2027

 Calibration Date Date Due

Certificate of Calibration

Customer: AIRFLOW PROS
1001 EASTWIND DRIVE SUITE 203
WESTERVILLE, OHIO 43081
614-807-5555

P.O. Number:

ID Number: 221200069



Description: TRMS CLAMP METER
Manufacturer: AMPROBE
Model Number: AMP-220
Serial Number: 221200069
Technician: LOGAN WATTS

Calibration Date: 11/17/2025
Calibration Due: 11/17/2026
Procedure: 33K1-4-2347-1
Rev: 8/30/2024
Temperature: 63 °F
Humidity: 31 % RH

As Found Condition: IN TOLERANCE
Calibration Results: IN TOLERANCE

On-Site Calibration:
Comments:

Limiting Attribute:

This instrument has been calibrated using standards traceable to the SI units through the National Institute of Standards and Technology (NIST) or other National Metrological Institute (NMI). The method of calibration is direct comparison to a known standard, derived from natural physical constants, ratio measurements or compared to consensus standards.

Reported uncertainties are expressed as expanded uncertainty values at an approximately 95% confidence level using a coverage factor of k=2. Statements of compliance are based on test results falling within specified limits with no reduction by the uncertainty of the measurement unless otherwise noted.


TMI's Quality System is accredited to ISO/IEC 17025:2017 and ANSI/ NCSL Z540-1-1994. ISO/IEC 17025:2017 is written in a language relevant to laboratory operations, meeting the principles of ISO 9001 and aligned with its pertinent requirements. This calibration complies with all the requirements of ANSI/ NCSL Z540-1-1994 and TMI's Quality Manual, QM-1.

Results contained in this document relate only to the item calibrated. Calibration due dates appearing on the certificate or label are determined by the client for administrative purposes and do not imply continued conformance to specifications.

This certificate shall not be reproduced, except in full, without the written permission of Technical Maintenance, Inc.

Measurements not currently on TMI's Scope of Accreditation are identified with an asterisk.


MATT AYRES, BRANCH MANAGER


Scott Chamberlain, QUALITY MANAGER

Calibration Standards

<u>Asset Number</u>	<u>Manufacturer</u>	<u>Model Number</u>	<u>Date Calibrated</u>	<u>Cal Due</u>
HLRD137	FLUKE	5522A	7/28/2025	7/28/2026
HLRD229	FLUKE	9100-200	2/7/2025	2/7/2027
HLRD664	HOBO	UX100-011A	10/14/2025	11/20/2026



Technical Maintenance, Inc.

4613 NORHTWEST PARKWAY, HILLIARD, OHIO 43026

ANSI/NCSL Z540-1-1994

Certificate of Calibration

Customer: AIRFLOW PROS
1001 EASTWIND DRIVE SUITE 203
WESTERVILLE, OHIO 43081
614-807-5555

P.O. Number:

ID Number: H.443221



Description: PHOTO/STROBE TACHOMETER
Manufacturer: EXTECH INSTRUMENTS
Model Number: 461825
Serial Number: H.443221
Technician: LOGAN WATTS

Calibration Date: 11/17/2025
Calibration Due: 11/17/2026
Procedure: 33K6-4-869-1
Rev: 11/30/2023
Temperature: 63 °F
Humidity: 31 % RH

As Found Condition: IN TOLERANCE
Calibration Results: IN TOLERANCE

On-Site Calibration:
Comments:

Limiting Attribute:

This instrument has been calibrated using standards traceable to the SI units through the National Institute of Standards and Technology (NIST) or other National Metrological Institute (NMI). The method of calibration is direct comparison to a known standard, derived from natural physical constants, ratio measurements or compared to consensus standards.

Reported uncertainties are expressed as expanded uncertainty values at an approximately 95% confidence level using a coverage factor of k=2. Statements of compliance are based on test results falling within specified limits with no reduction by the uncertainty of the measurement unless otherwise noted.

TMI's Quality System is accredited to ISO/IEC 17025:2017 and ANSI/ NCSL Z540-1-1994. ISO/IEC 17025:2017 is written in a language relevant to laboratory operations, meeting the principles of ISO 9001 and aligned with its pertinent requirements. This calibration complies with all the requirements of ANSI/ NCSL Z540-1-1994 and TMI's Quality Manual, QM-1.

Results contained in this document relate only to the item calibrated. Calibration due dates appearing on the certificate or label are determined by the client for administrative purposes and do not imply continued conformance to specifications.

This certificate shall not be reproduced, except in full, without the written permission of Technical Maintenance, Inc.

Measurements not currently on TMI's Scope of Accreditation are identified with an asterisk.

MATT AYRES, BRANCH MANAGER

Scott Chamberlain, QUALITY MANAGER

Calibration Standards

<u>Asset Number</u>	<u>Manufacturer</u>	<u>Model Number</u>	<u>Date Calibrated</u>	<u>Cal Due</u>
HLRD137	FLUKE	5522A	7/28/2025	7/28/2026
HLRD664	HOBO	UX100-011A	10/14/2025	11/20/2026



Technical Maintenance, Inc.

4613 NORHTWEST PARKWAY, HILLIARD, OHIO 43026

ANSI/NCSL Z540-1-1994