



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Water Pressure Sensing Module	Temperature	73	°F
Model	S-DP-250	Rel. Humidity	24	%
SN	2600108B	Bar. Pressure	28.8	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Specification +/- 2% +/- last digit	Cal Std	Sensor 1 Diff Pres	Difference	Cal Std	Sensor 2 Gage Lo	Diff
Approx Set Point			(%)			%
PSI	(ftwc)	(ftwc)		(ftwc)	(ftwc)	
0.13	0.3	0.3	0	0.3	0.3	0
0.434	1.0	1.0	0	1.0	1.0	0
0.867	2.0	2.0	0	2.0	2.0	0
2.168	5.0	5.0	0	5.0	5.0	0
	PSI	PSI		PSI	PSI	
10	10.0	10.0	0.0	10.0	10.0	0.0
40	40.0	40.1	0.3	40.0	40.1	0.3
80	80.0	80.4	0.5	80.0	80.4	0.5
120	120.0	120.0	0.0	120.0	120.0	0.0
240	240.0	240.1	0.0	240.0	240.1	0.0
Conversion						
PSI	ftwc	inwc				
1.00	2.307	27.68				
0.434	1.00	12.00				

Indicates out of tolerance condition -----↑

NIST-Traceable Calibration Lab Standards

Variable	System ID	Calibration Last	Calibration Due
Pressure	41001LGH	20-Jun-24	20-Jun-26
Pressure	41001JO3	19-Jun-24	19-Jun-26
Pressure	41001USY	21-Apr-25	21-Apr-27
Pressure	41001TJR	22-Apr-25	22-Apr-27

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

 Calibrated By

_____ 22-Jan-2026
Calibration Date
_____ 22-Jan-2027
Date Due



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Humidity Sensor	Temperature	75	°F
Model	PR-TH-12	Rel. Humidity	18	%
SN	2500123	Bar. Pressure	28.9	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Temperature (°F)	Spec				
	1	40.2	-1.0	1.0	40.1
	2	75.3	-1.0	1.0	75.4
	3	86.5	-1.0	1.0	86.3
	4	128.2	-2.0	2.0	128.4
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1	
	1	20.0			20.0
	2	28.7			28.7
	3	33.0			33.0
Humidity %RH 10 to 90%	Spec		-3	3	
	1	10.5			10.5
	2	21.8			22.4
	3	57.2			55.3
	4	90.1			90.8

Indicates out of tolerance condition -----↑

Calibration Standard

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	11-Aug-25	11-Aug-27
Temperature	21396189	24-Feb-25	24-Feb-27
Pressure	2205000006	10-Oct-25	10-Oct-27
Pressure	1208000080	14-Aug-25	14-Aug-27
Humidity	20558772	8-Oct-25	8-Oct-26
Humidity	20052171	13-Feb-25	13-Feb-26

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

Temperature accuracy (dry bulb) varies across the operating range:

Temperature over 32-100F +/- 1.0 F
 Temperature over 100-158F +/- 2.0 F

 Calibrated By

5-Feb-2026 5-Feb-2027
 Calibration Date Date Due

Customer: AIRFLOW PROS
1001 EASTWIND DRIVE SUITE 203
WESTERVILLE, OH 43081
614-807-5555

P.O. Number:

ID Number: H.437770



Description: PHOTO TACHOMETER/STROBOSCOPE
Manufacturer: EXTECH INSTRUMENTS
Model Number: 461825
Serial Number: H.437770
Technician: JODY CHESTER

Calibration Date: 02/19/2026
Calibration Due: 02/19/2027
Procedure: 33K3-4-593-1
Rev: 4/30/2025
Temperature: 76 °F
Humidity: 40 % RH

As Found Condition: IN TOLERANCE
Calibration Results: IN TOLERANCE

On-Site Calibration:
Comments:

Limiting Attribute:

This instrument has been calibrated using standards traceable to the SI units through the National Institute of Standards and Technology (NIST) or other National Metrological Institute (NMI). The method of calibration is direct comparison to a known standard, derived from natural physical constants, ratio measurements or compared to consensus standards.

Reported uncertainties are expressed as expanded uncertainty values at an approximately 95% confidence level using a coverage factor of k=2. Statements of compliance are based on test results falling within specified limits with no reduction by the uncertainty of the measurement unless otherwise noted.

TMI's Quality System is accredited to ISO/IEC 17025:2017 and ANSI/ NCSL Z540-1-1994. ISO/IEC 17025:2017 is written in a language relevant to laboratory operations, meeting the principles of ISO 9001 and aligned with its pertinent requirements. This calibration complies with all the requirements of ANSI/ NCSL Z540-1-1994 and TMI's Quality Manual, QM-1.

Results contained in this document relate only to the item calibrated. Calibration due dates appearing on the certificate or label are determined by the client for administrative purposes and do not imply continued conformance to specifications.

This certificate shall not be reproduced, except in full, without the written permission of Technical Maintenance, Inc.

Measurements not currently on TMI's Scope of Accreditation are identified with an asterisk.

MATT AYRES, BRANCH MANAGER

Scott Chamberlain, QUALITY MANAGER

Calibration Standards

Asset Number	Manufacturer	Model Number	Date Calibrated	Cal Due
HLRD100	FLUKE	5522A/SC1100	5/22/2025	5/22/2026
HLRD150	MONARCH INSTRUMENT	SPSR-IM	7/10/2025	7/28/2026
HLRD150-1	MONARCH INSTRUMENT	ROS-P	7/10/2025	7/28/2026
HLRD202	AGILENT	53131A	7/10/2025	2/20/2027
HLRD664	HOBO	UX100-011A	10/14/2025	11/20/2026



Technical Maintenance, Inc.

4613 NORTHWEST PARKWAY, HILLIARD, OH 43026

ANSI/NCSL Z540-1-1994

Customer: AIRFLOW PROS
1001 EASTWIND DRIVE SUITE 203
WESTERVILLE, OH 43081
614-807-5555

P.O. Number:

ID Number: 200501509



Description: CLAMP METER
Manufacturer: AMPROBE
Model Number: AMP-220
Serial Number: 200501509
Technician: JODY CHESTER

Calibration Date: 02/19/2026
Calibration Due: 02/19/2027
Procedure: 33K1-4-2347-1
Rev: 8/30/2024
Temperature: 76 °F
Humidity: 40 % RH

On-Site Calibration:
Comments:

As Found Condition: IN TOLERANCE
Calibration Results: IN TOLERANCE

Limiting Attribute:

This instrument has been calibrated using standards traceable to the SI units through the National Institute of Standards and Technology (NIST) or other National Metrological Institute (NMI). The method of calibration is direct comparison to a known standard, derived from natural physical constants, ratio measurements or compared to consensus standards.

Reported uncertainties are expressed as expanded uncertainty values at an approximately 95% confidence level using a coverage factor of k=2. Statements of compliance are based on test results falling within specified limits with no reduction by the uncertainty of the measurement unless otherwise noted.

TMI's Quality System is accredited to ISO/IEC 17025:2017 and ANSI/NCSL Z540-1-1994. ISO/IEC 17025:2017 is written in a language relevant to laboratory operations, meeting the principles of ISO 9001 and aligned with its pertinent requirements. This calibration complies with all the requirements of ANSI/NCSL Z540-1-1994 and TMI's Quality Manual, QM-1.

Results contained in this document relate only to the item calibrated. Calibration due dates appearing on the certificate or label are determined by the client for administrative purposes and do not imply continued conformance to specifications.

This certificate shall not be reproduced, except in full, without the written permission of Technical Maintenance, Inc.

Measurements not currently on TMI's Scope of Accreditation are identified with an asterisk.


MATT AYRES, BRANCH MANAGER


Scott Chamberlain, QUALITY MANAGER

Calibration Standards

<u>Asset Number</u>	<u>Manufacturer</u>	<u>Model Number</u>	<u>Date Calibrated</u>	<u>Cal Due</u>
HLRD100	FLUKE	5522A/SC1100	5/22/2025	5/22/2026
HLRD664	HOBO	UX100-011A	10/14/2025	11/20/2026



Technical Maintenance, Inc.

4613 NORTHWEST PARKWAY, HILLIARD, OH 43026

ANSI/NCSL Z540-1-1994



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry			Calibration Environment		
Temperature Product		Module	Probe	Temperature	75	°F
Model			PR-T-4-6	Rel. Humidity	24	%
SN			2300090	Bar. Pressure	28.9	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Cal Lab Module & Test Probe	Spec				
Temperature (°F)	1	75.5	-0.3	+0.3	75.5
	2	205.5	-2.6	+2.6	204.9
	3	-5.1	-1.6	+1.6	-4.6

Indicates out of tolerance condition -----↑

Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	11-Aug-25	11-Aug-26
Temperature	21396189	19-Jan-26	19-Jan-27

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

 Calibrated By

4-Feb-2026 4-Feb-2028
 Calibration Date Date Due



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Temperature Product	Meter Sensor	Temperature	75	°F
Model	MS-TH-1	Rel. Humidity	19	%
SN	2100140B	Bar. Pressure	28.9	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Cal Lab Probe & Test Module	Spec				
Temperature (°F)	1	74.4	-0.3	+0.3	74.4
	2	205.7	-2.6	+2.6	206.0
	3	-5.2	-1.6	+1.6	-5.4

Indicates out of tolerance condition -----↑

Calibration Standard SN & Dates

Variable	System ID	Calibration Last	Calibration Due
Temperature	16320239	11-Aug-25	11-Aug-26
Temperature	21396189	19-Jan-26	19-Jan-27

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self-calibrated techniques.

 Calibrated By

4-Feb-2026 4-Feb-2028
 Calibration Date Date Due



Certificate of Calibration

Airflow Pros

Manufacturer	Evergreen Telemetry	Calibration Environment		
Product	Pressure / Velocity Module	Temperature	75	°F
Model	S-PVF-1	Rel. Humidity	19	%
SN	2100464C	Bar. Pressure	28.9	in Hg

As Found
 As Left
 In Tolerance
 Out of Tolerance

Calibration Data

Measurement Variable	Test Point	Cal Standard	Allowable Range		Test Instrument
			Min	Max	
Barometric Pressure (in Hg)	Spec		-2% - 0.1	+ 2% + 0.1	
	1	20.0			20.1
	2	29.0			29.0
	3	33.0			33.1
Differential Pressure (in wc)	Spec		-2%-.001	+2%+.001	
	1	10.00			9.984
	2	2.000			1.989
	3	0.5000			0.4978
	4	0.0500			0.0496
	5	-10.00			-10.029
Via Pitot >> Velocity Pressure >> (inW.C. / FPM) -3% -7	6	-0.0500			-0.0499
	7	0.00069 / 105	-3% - 7	+3% + 7	104
	8	0.0158 / 503			502

Indicates out of tolerance condition -----↑

NIST-Traceable Lab Calibration Standards

Variable	System ID	Calibration Last	Calibration Due
Pressure	7481227 / 7568470	15-Dec-25	15-Dec-26
Pressure	7871917/ 7870754	16-Jan-26	16-Jan-27
Pressure	2205000006	10-Oct-25	10-Oct-26
Pressure	1208000080	16-Jan-26	16-Jan-27
Pressure	4100109L	28-Apr-25	28-Apr-26
Velocity	2500067	10-Sep-25	10-Sep-27
Velocity	2100190A	11-Aug-25	11-Aug-27

This instrument has been checked for accuracy, calibrated to manufacturer's specifications, and found to be within the specified tolerance unless otherwise stated. It has been calibrated using measurement standards traceable to the National Institute of Standards and Technology, or accepted intrinsic standards of measurement, or derived by the ratio type of self calibrated techniques.

 Calibrated By

4-Feb-2026 4-Feb-2028
 Calibration Date Date Due